



SLOVENSKA
AKREDITACIJA
SIST EN ISO/IEC 17025
LK-008



LOTRIČ
METROLOGY

CERTIFICATE OF CALIBRATION

Certificate No: 275-419-18-1

Page 1 of 22

CALIBRATION OF:

Sound Level Meter:	2238-4188-BZ7126	No: 2562568
Microphone:	4188	No: 2555350
Date of receipt:	2018-06-26	
Certificate No:	275-419-18-1	

CLIENT:

Politie Brabant Zuid-Oost
Repelweg 6
5261 SM VUGHT
Netherland

Order No: 00299

CALIBRATION CONDITIONS:

Preconditioning: 6 hours at 23 °C
Temperature: 23.0 °C ± 3°C
Air Pressure: 98.3 kPa ± 3 kPa
Rel. humidity: 48.4 %RH ± 20 %RH

PROCEDURE:

Calibration was performed according to the standards SIST IEC 60651 and 60804 type 1 and internal manual ML10N60.

RESULTS:

Results with uncertainty are stated from the page 2 onward.

Date of Calibration: 2018-06-27

Certificate issued: 2018-06-28

Franc Celestina
Calibration Technician



Digitally Signed by Franc
Celestina
Date: 28.06.2018 12:29

Luka Dolenc
Approved signatory

Digitally Signed by:
Luka Dolenc
Date: 2018.06.28 14:01:42 +02'00'

CERTIFICATE OF CALIBRATION

Certificate No.:275-420-18-1

Page 1 of 2

CALIBRATION OF:

Calibrator: Brüel & Kjær 4231 No: 2564379
Date of receipt: 2018-06-26 IEC Class: 1

CLIENT:

Politie Brabant Zuid-Oost
Repelweg 6
5261 SM VUGHT
Netherland

Order No: 00299

CALIBRATION CONDITIONS:

Preconditioning: 4 hours at 23 °C
Temperature: 23 °C
Air Pressure: 97.69 kPa
Rel. Humidity: 54 %RH

PROCEDURE:

Calibration was performed according to the standards SIST IEC60942 and internal manual ML10N60.


RESULTS:

Results with uncertainty are stated from the page 2 onwards.

Date of Calibration:2018-06-28

Certificate issued:2018-06-28


Franc Celestina
Calibration Technician


Luka Dolenc
Approved Signatory



Digitally Signed by Franc
Celestina
Date:28.06.2018 12:28

Digitally Signed by:
Luka Dolenc
Date: 2018.06.28 14:01:46 +02'00'

CERTIFIKAT O KALIBRACIJI / CERTIFICATE OF CALIBRATION

Certifikat St.: / Certificate No: IMS-88-18

Stran / Page 1 od/ of 4

MERILO / CALIBRATION OF:

Merilnik vrtljajev / Tachometer: **Tachometer**

No: 242295

Datum prejema / Date of receipt: 2018-06-26

Certifikat / Certificate No: IMS-88-18

NAROCNIK / CLIENT:

Politie Brabant Zuid-Oost
Repelweg 6
5261 SM VUGHT
Netherland

Narocilo St. / Order No: 00299

POGOJI OKOLJA / CALIBRATION CONDITIONS:

Priprava / Preconditioning: 6 hours at 23 °C
Temperatura: 23.0 °C ± 3°C
Zr. Tlak: 101.3 kPa ± 3 kPa
Relativna vlaga: 50.0 %RH ± 20 %RH

KALIBRACIJSKI POSTOPEK / PROCEDURE:

Kalibracija je bila izvedena skladno s specifikacijami proizvajalca in delovnih pogojev.
Calibration was performed according manufacturer's specification and operation requirements.

REZULTATI / RESULTS:

Rezultati in merilna negotovost so navedeni na strani 2. / Results with uncertainty are stated from the page 2 onward.

Datum Kalibracije / Date of Calibration: 2018-06-28

Certifikat izdan / Certificate issued: 2018-06-28



Franc Celestina
Calibration Technician



Digitally Signed by Franc
Celestina
Date: 28.06.2018 12:30



Luka Dolenc
Approved signatory

Digitally Signed by:
Luka Dolenc
Date: 2018.06.28 14:01:11 +02'00'

CERTIFIKAT O KALIBRACIJI / CERTIFICATE OF CALIBRATION

Certifikat St.: / Certificate No: IMS-89-18

Stran / Page 1 od/ of 4

MERILO / CALIBRATION OF:

Merilnik vrtljajev / Tachometer: **AVL Tachometer**

No: 4302

Datum prejema / Date of receipt: 2018-06-26

Certifikat / Certificate No: IMS-89-18

NAROCNIK / CLIENT:

Politie Brabant Zuid-Oost
Repelweg 6
5261 SM VUGHT
Netherland

Narocilo St. / Order No: 00299

POGOJI OKOLJA / CALIBRATION CONDITIONS:

Priprava / Preconditioning: 6 hours at 23 °C
Temperatura: 23.0 °C ± 3°C
Zr. Tlak: 101.3 ± 3 kPa
kPa
Relativna vlaga: 50.0 %RH ± 20 %RH

KALIBRACIJSKI POSTOPEK / PROCEDURE:

Kalibracija je bila izvedena skladno s specifikacijami proizvajalca in delovnih pogojev.
Calibration was performed according manufacturer's specification and operation requirements.

REZULTATI / RESULTS:


Rezultati in merilna negotovost so navedeni na strani 2. / Results with uncertainty are stated from the page 2 onward.

Datum Kalibracije / Date of Calibration: 2018-06-28

Certifikat izdan / Certificate issued: 2018-06-28


Franc Celestina
Calibration Technician

Digitally Signed by Franc
Celestina
Date: 28.06.2018 12:30


Luka Dolenc
Approved signatory

Digitally Signed by:
Luka Dolenc
Date: 2018.06.28 14:01:16 +02'00'

