



SLOVENSKA  
AKREDITACIJA  
SIST EN ISO/IEC 17025  
LK-008



LOTRIČ<sup>®</sup>  
METROLOGY

## CERTIFICATE OF CALIBRATION

Certificate No: 275-61-15-1

Page 1 of 13

### CALIBRATION OF:

Sound Level Meter:	2238-4188-BZ7126	No: 2562568
Microphone:	4188	No: 2555350
Date of receipt:	2015-02-26	
Certificate No:	275-61-15-1	

### CLIENT:

Politie Brabant Zuid-Oost  
Repelweg 6  
5261 SM VUGHT  
Netherland

Order No: 00083

### CALIBRATION CONDITIONS:

Preconditioning:: 6 hours at 23 °C  
Temperature: 22.9 °C ± 3°C  
Air Pressure: 97.7 kPa ± 3 kPa  
Rel. humidity: 32.0 %RH ± 20 %RH

### PROCEDURE:

Calibration was performed according to the standards SIST IEC 60651 and 60804 type 1 and internal manual ML10N60.

### RESULTS:

Results with uncertainty are stated from the page 2 onward.

Date of Calibration: 2015-03-02

Certificate issued: 2015-03-02

Franc Celestina  
Calibration Technician

Digitally Signed by Franc  
Celestina  
Date: 03.03.2015 13:06:48



Luka Dolenc  
Approved signatory

Digitally Signed by Luka  
Dolenc  
Date: 2015.03.03 14:12:41  
+0100

## CERTIFIKAT O KALIBRACIJI / CERTIFICATE OF CALIBRATION

Certifikat St.: / Certificate No: IMS-22-15

Stran / Page 1 od/ of 4

### MERILO / CALIBRATION OF:

Merilnik vrtiljajev / Tachometer: Tachometer

No: 242295

Datum prejema / Date of receipt: 2015-02-26

Certifikat / Certificate No: IMS-22-15

### NAROCNIK / CLIENT:

Politie Brabant Zuid-Oost  
Repelweg 6  
5261 SM VUGHT  
Netherland

Narocilo St. / Order No: 00083

### POGOJI OKOLJA / CALIBRATION CONDITIONS:

Priprava / Preconditioning: 6 hours at 23 °C  
Temperatura: 23.0 °C ± 3°C  
Zr. Tlak: 97.5 kPa ± 3 kPa  
Relativna vlaga: 32.2 %RH ± 20 %RH

### KALIBRACIJSKI POSTOPEK / PROCEDURE:

Kalibracija je bila izvedena skladno s specifikacijami proizvajalca in delovnih pogojev.  
Calibration was performed according manufacturer's specification and operation requirements.

### REZULTATI / RESULTS:

Rezultati in merilna negotovost so navedeni na strani 2. / Results with uncertainty are stated from the page 2 onward.

Datum Kalibracije / Date of Calibration: 2015-03-02

Certifikat izdan / Certificate issued: 2015-03-02



Franc Celestina  
Calibration Technician

Digitally Signed by Franc  
Celestina  
Date: 03.03.2015 13:07:01



Digitally Signed by Luka  
Dolenc  
Date: 2015.03.03 14:13:15  
+0100



Luka Dolenc  
Approved signatory



SLOVENSKA  
AKREDITACIJA  
SIST EN ISO/IEC 17025  
LK-008



LOTRIČ  
METROLOGY

## CERTIFICATE OF CALIBRATION

Certificate No.:275-60-15-1

Page 1 of 3

### CALIBRATION OF:

Calibrator: Brüel & Kjær 4231 No: 2564379  
Date of receipt: 2015-02-26 IEC Class: 1

### CLIENT:

Politie Brabant Zuid-Oost  
Repelweg 6  
5261 SM VUGHT  
Netherland

Order No: 00083

### CALIBRATION CONDITIONS:

Preconditioning: 4 hours at 23 °C  
Temperature: 23.02 °C  
Air Pressure: 97.71 kPa  
Rel. Humidity: 31.5 %RH

### PROCEDURE:

Calibration was performed according to the standards SIST IEC60942 and internal manual ML10N60.

### RESULTS:


Results with uncertainty are stated from the page 2 onwards.

Date of Calibration:2015-03-02

Certificate issued:2015-03-02

Franc Celestina  
Calibration Technician

Luka Dolenc  
Approved Signatory



Digitally Signed byFranc  
Celestina  
Date:03.03.2015 13:06:17

Digitally Signed by Luka  
Dolenc  
Date: 2015.03.03 14:12:28  
+0100